Device Description	on:		
Contact Name:		Phone:	
Company:		Address:	
E-mail:		City:	
	Instructions For Completing	Pre-Evaluation Checklists	;
most applicable to system the "General be completed, submitted with an The exceptions a have the "General beautiful	need to complete both the "Gene to your device or system type. Feral" checklist (which applies to al Both the "General" and "Watthon electric watthour metering system are the computer software/hardwal" requirements incorporated in	For example, for a weighing I device types) and the "Scaur Meter" checklists should application. Fare component pre-evaluate them. Use the checklist for	d device or weighing ales" checklist should d be completed and tion checklists which or computer systems
	either scales or measuring syste e software will be connected to bo		ion checklist will be
not all-encompas single device typ	include requirements extracted fresing, the checklists contain require or accessory. It is best to thin as a component of such a system	rements beyond those which ik of a device type as a we	h would apply to any eighing or measuring
When applying th	ne requirements to your device you	u have three options; Check	
YES NO NA	If your device or system complie if the device or system does not if sections appear not to apply to	comply.	s)
such a nature that	consider if your device or system at it will not effect the ability to te nents or lack of provision for se being corrected.	est for accuracy, such as fai	ilure to conform with
If you are not abl an evaluation.	le to conduct accuracy testing you	ır system or device is probe	ably not yet ready for
which we have	the enclosed specifications, tole applied for evaluation and applevice meets all applicable require	roval. To the best of my	- .
Signed:		Date:	

A. Application.

- **A.1.** This code applies to devices used for the measurement of hydrocarbon gas in the vapor state, such as propane, propylene, butanes, butylenes, ethane, methane, natural gas and any other hydrocarbon gas/air mix.
- **A.2.** This code does not apply to:
- (a) Liquid-measuring devices used for dispensing liquefied petroleum gases in liquid form.
- (b) Natural, liquefied petroleum, and manufactured-gas-vapor meters when these are operated in a public utility system.
- (c) Mass flow meters.
- A.3. See also General Code requirements.

	Yes	No	NA
S. Specifications.			
S.1. Design of Indicating and Recording Elements and of Recorded Representations.			
S.1.1. Primary Elements.			
S.1.1.1. General. - A device shall be equipped with a primary indicating element and may also be equipped with a primary recording element.			
S.1.1.2. Units. - A volume-measuring device shall indicate, and record if equipped to record, its deliveries in terms of cubic meters or cubic feet, or multiple or decimal subdivisions of cubic meters or cubic feet.			
S.1.1.3. Value of Smallest Unit Volume-Measuring Devices: The value of the smallest unit of indicated delivery, and recorded delivery if the device is equipped to record, shall not exceed:			
 (a) 1 m³ (1 000 dm³) (100 ft³) when the maximum rated gas capacity is less than 100 m³/h (10 000 ft³/h); (b) 10 m³ (1 000 ft³) when the maximum rated gas capacity is 280 m³/h (10 000 ft³/h) up to but not including 1 700 m³/h (60 000 ft³/h); (c) 100 m³ (10 000 ft³) when the maximum rated gas capacity is 1 700 m³/h (60 000 ft³/h) or more. 			
S.1.1.4. Advancement of Indicating and Recording Elements Primary indicating and recording elements shall advance digitally or continuously and be susceptible to advancement only by the mechanical operation of the device.			

	Yes	No	NA
S.1.1.5. Proving Indicator. - Devices rated less than 280 m³/h (10 000 ft³/h) gas capacity shall be equipped with a proving indicator measuring 0.025, 0.05, 0.1, 0.2, or 0.25 m³ per revolution (1, 2, 5, or 10 ft³ per revolution) for testing the meter. Devices with larger capacities shall be equipped as follows:			
 (a) Devices rated 280 m³ (10 000 ft³) up to but not including 1 700 m³/h (60 000 ft³/h) gas capacity shall be equipped with a proving indicator measuring not greater than 1 m³ (100 ft³) per revolution. (b) Devices rated 1 700 m³/h (60 000 ft³/h) gas capacity or more shall be equipped with a proving indicator measuring not more than 10 m³ (1 000 ft³) per revolution. 			
The test circle of the proving indicator shall be divided into 10 equal parts. Additional subdivisions of one or more of such equal parts may be made.			
S.1.2. Graduations.			
S.1.2.1. Length. - Graduations shall be so varied in length that they may be conveniently read.			
S.1.2.2. Width. - In any series of graduations, the width of a graduation shall in no case be greater than the width of the minimum clear interval between graduations, and in no case should it exceed 1.0 mm (0.04 in) for indicating elements and 0.5 mm (0.02 in) for proving circles.			
S.1.2.3. Clear Interval Between Graduations The clear interval shall be not less than 1.0 mm (0.04 in). If the graduations are not parallel, the measurement shall be made:			
(a) along the line of relative movement between the graduations at the end of the indicator,			
or (b) if the indicator is continuous, at the point of widest separation of the graduations.			
S.1.3. Indicators.			
S.1.3.1. Symmetry. - The index of an indicator shall be symmetrical with respect to the graduations, at least throughout that portion of its length associated with the graduations.			
S.1.3.2. Length. - The index of an indicator shall reach to the finest graduations with which it is used.			
S.1.3.3. Width. - The width of the index of an indicator in relation to the series of graduations with which it is used shall be not greater than:			
(a) the width of the widest graduation, and(b) the width of the minimum clear interval between graduations.			
When the index of an indicator extends along the entire length of a graduation, that portion of the index of the indicator that may be brought into coincidence with the graduation shall be of the same width throughout the length of the index that coincides with the graduation.			

	Yes	No	NA
S.1.3.4. Clearance The clearance between the index of an indicator and the graduations shall in no case be more than 1.5 mm (0.06 in).			
S.1.3.5. Parallax Parallax effects shall be reduced to the practicable minimum.			
S.2. Design of Measuring Elements.			
S.2.1. Pressure Regulation. - Except when measured as a retail motor fuel, the vapor should be measured at a normal gauge pressure (psig) of:			
 (a) 2 740 Pa 685 Pa [11 in of water column (0.40 psig) 2.75 in of water column (0.10 psig)] for liquefied petroleum gas vapor; or (b) 1 744 Pa 436 Pa [7 in of water column (0.25 psig) 1.75 in of water column (0.06 psig)] for natural and manufactured gas. 			
When vapor is measured at a pressure other than what is specified above for the specific product, a volume multiplier shall be applied within the meter or to the billing invoice based on the following equation:			
$VPM = \frac{AAP + GP}{AAP + NGP}$			
Where			
VPM = Volume pressure multiplier AAP = Assumed atmospheric pressure in psia GP = Gauge pressure in pascal or psig NGP = Normal gauge pressure in pascal or psig			
The assumed atmospheric pressure is to be taken from Tables 2 and 2M (see pages 7 & 8).			
When liquefied petroleum gas vapor is measured at a pressure of 6 900 Pa (1 psig) or more, the delivery pressure shall be maintained within 1 725 Pa (0.25 psig).			
Pressure variations due to regulator lock off shall not increase the operating pressure by more than 25%.			
S.2.2. Provision for Sealing. - Adequate provision shall be made for applying security seals in such a manner that no adjustment may be made of any measurement element.			
S.2.3. Maintenance of Vapor State. - A device shall be so designed and installed that the product being measured will remain in a vapor state during passage through the meter.			
S.2.4. Automatic Temperature Compensation. - A device may be equipped with an adjustable automatic means for adjusting the indication and registration of the measured volume of vapor product to the volume at 15 °C (60 °F).			

	Yes	No	NA
S.3. Design of Discharge Lines.			
S.3.1. Diversion of Measured Vapor. - No means shall be provided by which any measured vapor can be diverted from the measuring chamber of the meter or the discharge line therefrom.			
S.4. Marking Requirements.			
S.4.1. Limitations of Use If a device is intended to measure accurately only products having particular properties, or to measure accurately only under specific installation or operating conditions, or to measure accurately only when used in conjunction with specific accessory equipment, these limitations shall be clearly and permanently stated on the device.			
S.4.2. Discharge Rates. - A volume-measuring device shall be marked to show its rated gas capacity in cubic meters per hour or cubic feet per hour.			
S.4.3. Temperature Compensation. [NOT ADOPTED]			
S.4.4. Badge. - A badge affixed in a prominent position on the front of the device shall show the manufacturer's name, serial number and model number of the device, and capacity rate of the device for the particular products that it was designed to meter as recommended by the manufacturer.			
N. Notes.			
N.1. Test Medium The device shall be tested with air or the product to be measured.			
N.2. Temperature and Volume Change Care should be exercised to reduce to a minimum any volume changes. The temperature of the air, bell-prover oil, and the meters under test should be within 1 °C (2 °F) of one another. The devices should remain in the proving room for at least 16 hours before starting any proving operations to allow the device temperature to approximate the temperature of the proving device.			
N.3. Test Drafts. - Except for low-flame tests, test drafts shall be at least equal to one complete revolution of the largest capacity proving indicator, and shall in no case be less than 0.05 m ³ or 2 ft ³ . All flow rates shall be controlled by suitable outlet orifices.			
N.4. Test Procedures. - If a device is equipped with an automatic temperature compensator, the proving device reading shall be corrected to 15 °C (60 °F), using an approved table.			
N.4.1. Normal Tests. - The normal test of a device shall be made at a rate not to exceed the capacity rate given on the badge of the meter.			
N.4.1.1. Automatic Temperature Compensation. - If a device is equipped with an automatic temperature compensator, the quantity of the test draft indication of the standard shall be corrected to 15°C (60 °F).			
N.4.2. Special Tests. - "Special" tests, to develop the operating characteristics of a device, and any special elements and accessories attached to or associated with the device, shall be made as circumstances require. Any test except as set forth in N.4.1. is a special test.			

	Yes	No	NA
N.4.2.1. Slow Test. - The device shall be tested at a rate not less than 20 percent of the marked capacity rate, or (at the check rate) not less than the minimum flow rate if marked on the device, whichever is less.			
N.4.2.2. Low-Flame Test. - The device shall be tested at an extremely low-flow rate as given in Table 1 (see page 7). The test shall consist of passing air at a pressure of 375 Pa (1.5 in water column) through the meter for not less than 60 minutes. The meter shall continue to advance at the conclusion of the test period.			
N.4.2.3. Pressure Regulation Test. - On devices operating at a pressure of 6 900 Pa (1 psig) or more, a pressure regulation test shall be made at both the minimum and maximum use load to determine the proper operation of the regulator and the proper sizing of the piping and dispensing equipment. These tests may include a test of 24 hours during which the pressure is recorded.			
N.5. Temperature Correction. - Corrections shall be made for any changes in volume resulting from the difference in air temperatures between time of passage through the device and time of volumetric determination in the proving device.			
 N.6. Frequency of Test A hydrocarbon gas vapor-measuring device shall be tested before installation and allowed to remain in service for 10 years from the time last tested without being retested, unless a test is requested by: (a) the purchaser of the product being metered, (b) the seller of the product being metered, or (c) the weights and measures official. 			
T. Tolerances.			
T.1. Tolerance Values on Normal Tests and on Special Tests Other Than Low-Flame Tests Maintenance and acceptance tolerances for hydrocarbon gas vapor-measuring devices shall be 3 percent (1.03 proof) of the test draft on underregistration and 1.5 percent (0.985 proof) of the test draft on overregistration.			

Table 1 Capacity of Low-Flow Test Rate Orifices With Respect to Device Capacity								
Metric Ur	nits	Inch-Pound U	nits					
Rated Capacity	Low-Flow Test Rate	Rated Capacity	Low-Flow Test Rate					
Up to and including 7 m ³ /h	0.007 m ³ /h	Up to and including 250 ft ³ /h	0.25 ft ³ /h					
Over 7 m ³ /h up to and including 14 m ³ /h	0.014 m ³ /h	Over 250 ft ³ /h up to and including 500 ft ³ /h	0.50 ft ³ /h					
Over 14 m³/h	0.1% of capacity rate	Over 500 ft ³ /h	0.1% of capacity rate					

	Table 2 Corrections for Altitude, Inch-Pound Units								
Elevation			Corre			Pressure I	ed Atmospheric Ire Plus Gauge Pressure		
	(feet)		11 inch WC	7 inch WC	(psia)	11 inch WC (psia)	7 inch WC (psia)		
	-150 to	400	1.02	1.01	14.64	15.04	14.89		
above	400 to	950	1.00	0.99	14.35	14.74	14.60		
above	950 to	1550	0.98	0.97	14.05	14.45	14.30		
above	1550 to	2100	0.96	0.95	13.76	14.15	14.01		
above	2100 to	2700	0.94	0.93	13.46	13.86	13.71		
above	2700 to	3300	0.92	0.91	13.17	13.56	13.42		
above	3300 to	3950	0.90	0.89	12.87 12.58	13.27	13.12		
above	3950 to 4550 to	4550 5200	0.88 0.86	0.87 0.85	12.58	12.97 12.68	12.83 12.53		
above above	5200 to	5200 5850	0.84	0.83	12.26	12.88	12.53		
above	5850 to	6500	0.82	0.83	11.69	12.36	11.94		
above	6500 to	7200	0.80	0.79	11.40	11.79	11.65		
above	7200 to	7900	0.78	0.73	11.10	11.50	11.35		
above	7900 to	8600	0.76	0.77	10.81	11.20	11.06		
above	8600 to	9350	0.74	0.73	10.51	10.91	10.76		
above	9350 to	10 100	0.72	0.71	10.22	10.61	10.47		
above	10 100 to	10 850	0.70	0.69	9.92	10.32	10.17		
above	10 850 to	11 650	0.68	0.67	9.63	10.03	9.88		
above	11 650 to	12 450	0.66	0.65	9.33	9.73	9.58		
above	12 450 to	13 250	0.64	0.63	9.04	9.44	9.29		
above	13 250 to	14 100	0.62	0.61	8.75	9.14	9.00		
above	14 100 to	14 950	0.60	0.59	8.45	8.85	8.70		

Table 2M Corrections for Altitude, Metric Units									
Elevation			Correction Atmospheri Pres		Pressu	Atmospheric sure Plus Pressure			
	(meters)		2.74 kPa Gauge Pressure	1.74 kPa Gauge Pressure	(kPa)	2.74 kPa 1.74 kPa Gauge Gauge Pressure Pressure			
above	-50 to 120 to 300 to 470 to 650 to 830 to 1020 to 1210 to 1400 to 1590 to 1790 to 2000 to 2210 to 2420 to 2640 to 2860 to 3080 to 3320 to 3560 to 3800 to 4050 to	120 300 470 650 830 1020 1210 1400 1590 1790 2000 2210 2420 2640 2860 3080 3320 3560 3800 4050 4310	1.02 1.00 0.98 0.96 0.94 0.92 0.90 0.88 0.86 0.84 0.82 0.78 0.76 0.74 0.72 0.70 0.68 0.66 0.64 0.62	1.01 0.99 0.97 0.95 0.93 0.91 0.89 0.87 0.85 0.83 0.81 0.79 0.77 0.75 0.73 0.71 0.69 0.67 0.65 0.63	100.85 98.82 96.79 94.76 92.73 90.70 88.66 86.63 84.60 82.57 80.54 78.51 76.48 74.45 72.41 70.38 68.35 66.32 64.29 62.26 60.23	103.59 101.56 99.53 97.50 95.47 93.44 91.40 89.37 87.34 85.31 83.28 81.25 79.22 77.19 75.15 73.12 71.09 69.06 67.03 65.00 62.97	102.58 100.54 98.51 96.48 94.45 92.42 90.39 88.36 86.33 84.29 82.26 80.23 78.20 76.17 74.15 72.12 70.08 68.05 66.01 63.98 61.95		